


2.50 eq.



Current Data Parameters
NAME 2022-09-19-RS-401
EXPNO 9
PROCNO 1

F2 - Acquisition Parameters
Date_ 20220919
Time 9.51 h
INSTRUM AVIII400NB
PROBHD Z108618_0022 (
PULPROG zg30
TD 65536
SOLVENT CDCl3
NS 16
DS 2
SWH 8000.000 Hz
FIDRES 0.244141 Hz
AQ 4.0960002 sec
RG 372.5
DW 62.500 usec
DE 10.52 usec
TE 298.0 K
D1 1.00000000 sec
TD0 1
SFO1 400.1324710 MHz
NUC1 1H
P0 4.40 usec
P1 13.20 usec
PLW1 20.00000000 W

F2 - Processing parameters
SI 65536
SF 400.1300000 MHz
WDW EM
SSB 0
LB 0.30 Hz
GB 0
PC 2.00

