

0.6 eq (506 uL in total)



Current Data Parameters
NAME 2023-06-01-SB137 Na Titration
EXPNO 4
PROCNO 1

F2 - Acquisition Parameters
Date_ 20230601
Time 9.46 h
INSTRUM AVIII400NB
PROBHD Z108618_0022 (
PULPROG zg30
TD 65536
SOLVENT CD3CN+CDCl3
NS 16
DS 2
SWH 8000.000 Hz
FIDRES 0.244141 Hz
AQ 4.0960002 sec
RG 400.79
DW 62.500 usec
DE 10.77 usec
TE 298.0 K
D1 1.00000000 sec
TD0 1
SF01 400.1324710 MHz
NUC1 1H
P0 4.30 usec
P1 12.90 usec
PLW1 20.00000000 W

F2 - Processing parameters
SI 65536
SF 400.1300000 MHz
WDW EM
SSB 0
LB 0.30 Hz
GB 0
PC 2.00

