



Current Data Parameters
NAME 2022-01-14-SB80 (Full Data)
EXPNO 3
PROCNO 1

F2 - Acquisition Parameters
Date_ 20220115
Time 1.51 h
INSTRUM AVIII400NB
PROBHD Z108618_0022 (
PULPROG hsqcetdgppsp.3
TD 2048
SOLVENT MeOD-d4+CDCl3
NS 2
DS 16
SWH 4795.396 Hz
FIDRES 4.683004 Hz
AQ 0.2135381 sec
RG 61.79
DW 104.267 usec
DE 13.72 usec
TE 298.1 K
CNST2 145.0000000
D0 0.00000300 sec
D1 1.00000000 sec
D4 0.00172414 sec
D11 0.03000000 sec
D16 0.00020000 sec
D21 0.00344828 sec
IN0 0.00003010 sec
TDav 1
ZGPGTNS
SFO1 400.1318806 MHz
NUC1 1H
P1 12.15 usec
P2 24.30 usec
PLW1 20.00000000 W
SFO2 100.6203150 MHz
NUC2 13C
CPDPRG[2] garp4
P3 9.35 usec
P14 500.00 usec
P31 2119.00 usec
PCPD2 80.00 usec
PLW0 0 W
PLW2 75.00000000 W
PLW12 1.02450001 W
SPNAM[3] Crp60,0.5,20.1
SPOAL3 0.500
SPOFFS3 0 Hz
SPW3 10.01799965 W
SPNAM[18] Crp60_xfilt.2
SPOAL18 0.500
SPOFFS18 0 Hz
SPW18 1.92990005 W
GPNAM[1] SMSQ10.100
GPZ1 80.00 %
GPNAM[2] SMSQ10.100
GPZ2 20.10 %
P16 1000.00 usec

F1 - Acquisition parameters
TD 256
SFO1 100.6203 MHz
FIDRES 129.775742 Hz
SW 165.089 ppm
FnMODE Echo-Antiecho

F2 - Processing parameters
SI 1024
SF 400.1300000 MHz
WDW QSINE
SBB 2
LB 0 Hz
GB 0
PC 1.40

F1 - Processing parameters
SI 1024
MC2 echo-antiecho
SF 100.6127690 MHz
WDW QSINE
SBB 2
LB 0 Hz
GB 0

